



Department of Facilities POLICY AND PROCEDURE

POLICY NUMBER: 268
VERSION: 1

SUBJECT: FIXED MEDICAL EQUIPMENT (HARD WIRED)

POLICY:

All hard wired equipment will be inspected and serviced at least once a year or as per Regulatory Compliance, The Joint Commission Standards, the Preventative Maintenance Program and manufactures specifications.

PURPOSE:

To ensure the safe and proper operation of all hard wired equipment with in the Facility.

PROCEDURE:

All hard wired equipment will be inspected at least once a year unless otherwise specified. This inspection will include checking that ground wire is fixed and tight, and the equipment has a ground resistance reading of NOT more than 500 milli ohms.

Approved By: Anthony Corliss (MANAGER I, FACILITIES OPERATIONS AND CRAFTS)	
Date: 10/11/2013	Original Date: 01/04/1994
Reviewed: 04/12/2016	Next Review Date: 04/12/2017
Supersedes:	