

VALLEYCARE
OLIVE VIEW-UCLA MEDICAL CENTER/HEALTH CENTERS
DEPARTMENT: FACILITIES
POLICY & PROCEDURE

NUMBER: 574
VERSION: 1

SUBJECT/TITLE: LABORATORY EQUIPMENT

POLICY: All laboratory equipment will be tested at intervals not to exceed six (6) months.

PURPOSE:

DEPARTMENTS: FACILITIES

DEFINITIONS:

PROCEDURE: An electrical safety test will be conducted using a current leakage meter and a visual test of all three wire cords and caps. A functional test will be made by operating the equipment to insure it is in good operating condition. Document all tests in the logs.

References:	
Approved by: Robert Ross (Director, Facilities)	Date: 08/12/2009
Review Date: 07/12/2015	Revision Date:
Distribution: Facilities	
Original Date: 08/12/2009	