

VALLEYCARE
OLIVE VIEW-UCLA MEDICAL CENTER/HEALTH CENTERS
DEPARTMENT: FACILITIES
POLICY & PROCEDURE

NUMBER: 614
VERSION: 1

SUBJECT/TITLE: ELECTRIC BEDS

POLICY: All electric beds will be electrical safety and functional tested at intervals not to exceed six (6) months.

PURPOSE:

PROCEDURE: Test for electric beds will be conducted with a current leakage meter and leakage shall be no more than 100 micro amps. All beds will have a three (3) wire cord and hospital grade cap. Visual inspection of cord and cap will be made. Operate bed in all modes to insure proper operation. Document all tests on P.M. sheets.

DEPARTMENTS: FACILITIES

DEFINITIONS:

References:	
Approved by: Robert Ross (Director, Facilities)	Date: 08/12/2009
Review Date: 07/12/2015	Revision Date:
Distribution: Facilities	
Original Date: 08/12/2009	